



Food and Drug Administration 2098 Gaither Road Rockville MD 20850

JUL 2 8 2005

James A. Bourland, Ph.D.
Technical Director
Quest Diagnostics Incorporated
4230 Burnham Avenue
Las Vegas, NV 89119

Re:

k042726

Trade/Device Name: Quest Diagnostics HairCheck-DT (PCP)

Regulatory Class: Class II Product Code: LCM Dated: April 4, 2005

Received: May 4, 2005

Dear Dr. Bourland:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820).

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific information about the application of labeling requirements to your device, or questions on the promotion and advertising of your device, please contact the Office of In Vitro Diagnostic Device Evaluation and Safety at (240) 276-0484. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/industry/support/index.html.

Sincerely yours,

Carol C. Benson, M.A.

Acting Director

Division of Chemistry and Toxicology

Office of In Vitro Diagnostic Device

Carol C. Benson

Evaluation and Safety

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510 (k) Number: k042726

Device Name: Quest Diagnostics HairCheck-DT (PCP)

Indications for use:

1. Intended Use and Indications for Use of the Subject Device:

The Quest Diagnostics Hair Check-DT (PCP) is a test system that utilizes the IDS One-Step ELISA PCP Kit for the qualitative detection of Phencyclidine at or above 300 pg/mg in head hair samples. This test system has not been evaluated for use in specific user populations or with hair specimens other than the head. It is an in vitro diagnostic device intended exclusively for in-house professional use only and is not intended for sale to anyone.

The Quest Diagnostics Hair Check-DT (PCP) screening test provides only a preliminary analytical test result. A more specific alternate chemical method must be used in order to obtain a confirmed result. Gas Chromatograph - Mass Spectrometry operating in the selected ion monitoring (SIM) mode is the preferred method with deuterated internal standards. Other chemical confirmation methods are available. Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are obtained.

Prescription Use(Part 21 CFR 801 Subpart D)	AND/OR	Over-The-Counter Use (21 CFR 807 Subpart C)
(PLEASE DO NOT WRITE BE NEEDED)	ELOW THIS LINE-	CONTINUE ON ANOTHER PAGE IF
		ALC:
Concurrence of CDR	H, Office of In Vitr	o Diagnostic Devices (OIVD)

Division Sign-Off

Office of In Vitro Diagnostic
Device Evaluation and Safety

Page 1 of ____

510Ki 6042726